



Republikang Pilipinas
Kagawaran ng Kalusugan
Panlabas ng Pagpapaganap
Pampurok ng Tanggapan ng Pangkalusugan Bldg. 6
WESTERN VISAYAS MEDICAL CENTER
Q. Abeto St., Mandurriao, Iloilo City
Tel. No.: (033) 321-2841 Telefax no.: (033) 321-1797
wvmciloilo@gmail.com



BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL/BID BULLETIN

Addendum No. 2

(Procurement of Various Laboratory Supplies, Reagents and Consumables
with Free Use of Machine for CY 2019)

This Addendum No. 2 dated March 1, 2019 is issued to amend items in the Technical Specifications for the aforementioned project on March 6, 2019.

Technical Specifications

ORIGINAL

I. LINE ITEMS

Item #	Item Description	Unit	Quantity	ABC/Unit	Total ABC	Statement of Compliance
Chemistry						
44	75g OGTT Bottle	Bottle	150	150.00	22,500.00	
45	50g OGTT Bottle	Bottle	50	150.00	7,500.00	
46	100g OGTT Bottle	Bottle	50	150.00	7,500.00	
46	Serum Separator plastic tube 4ml Z serum Clot activator, 13x75mm tube, redtop with yellow ring, 100's/pack	Pack	3500	850.00	2,975,000.00	

II. Lot Items

Lot #	Item Description	Unit	Quantity	ABC/Unit	Total ABC	Statement of Compliance
2	Specific Protein Analyzer Specifications: Principle: Nephelometry Throughput: 160 samples /hour Sample Volume : 2-20 ul Quantitatively analyze the complete parameters of the following: *HS-CRP *C-Reactive Protein * Anti-Streptolysin O *Rheumatoid Factor *C3 and C4 * D-Dimer Calibration by magnetic card With 4 independent channels Reaction Time: 1.5 - 3 minutes Display: 5.6 inches color touch screen	Per test Per test Per test Per test Per test Per test	3000 8000 300 300 1000 1200		3,559,000.00	

	Storage: 100,000 results storage Disposable cuvette, no carry over Auto and manual printing thermal paper Working Temperature 10-30°C Support LIS and barcode, WEBLIS compatible Less and easy maintenance Reagents including Controls: Provide 1 bottle of Normal and pathological Control for every 100 Tests ASO C-Reactive Protein HS- C-Reactive Protein Rheumatoid Factor C3 D-Dimer					
--	--	--	--	--	--	--

AMENDED

I. LINE ITEMS

Item #	Item Description	Unit	Quantity	ABC/Unit	Total ABC	Statement of Compliance
	Chemistry					
44	75g OGTT Bottle	Bottle	150	150.00	22,500.00	
45	50g OGTT Bottle	Bottle	50	150.00	7,500.00	
46	100g OGTT Bottle	Bottle	50	150.00	7,500.00	
46A	Serum Separator plastic tube 4ml Z serum Clot activator, 13x75mm tube, redtop with yellow ring, 100's/pack	Pack	3500	850.00	2,975,000.00	

II. Lot Items

Lot #	Item Description	Unit	Quantity	ABC/Unit	Total ABC	Statement of Compliance
2	Specific Protein Analyzer Specifications: <u>Fully automated</u> Principle: Nephelometry Throughput: 160 samples /hour Sample Volume : 2-20 ul Quantitatively analyze the complete parameters of the following: *HS-CRP * C-Reactive Protein * Anti-Streptolysin O * Rheumatoid Factor * C3 and C4 * D-Dimer Calibration by magnetic card <u>With 4 independent channels</u> Reaction Time: 1.5 - 3 minutes <u>Display: 5.6 inches color touch screen</u> Storage: 100,000 results storage Disposable cuvette, no carry over Auto and manual printing thermal paper Working Temperature 10-30°C Support LIS and barcode, WEBLIS compatible Less and easy maintenance Reagents including Controls: Provide 1 bottle of Normal and pathological Control for every 100 Tests ASO C-Reactive Protein				3,559,000.00	
		<u>Per test</u>	3000			
		<u>Per test</u>	8000			

HS- C-Reactive Protein	<u>Per test</u>	<u>300</u>			
Rheumatoid Factor	<u>Per test</u>	<u>300</u>			
C3	<u>Per test</u>	<u>1000</u>			
D-Dimer	<u>Per test</u>	<u>1200</u>			
All other items in the Technical Specifications remain the same.					

This shall form an integral part of the Bid Documents.

For guidance and information of all concerned:

(SGD.) GLENN R. MACLANG, MD
Vice-Chairman, Bids and Awards Committee