



Republika ng Pilipinas  
Kagawaran ng Kalusugan  
Panlabas ng Pagpapaganap  
Pampurok ng Tanggapang Pangkalusugan Blg. 6  
**WESTERN VISAYAS MEDICAL CENTER**  
Q. Abeto St., Mandurriao, Iloilo City  
Tel. No.; (033) 321-2841 Telefax no.: (033) 321-1797  
[wvmciloilo@gmail.com](mailto:wvmciloilo@gmail.com)



## **BIDS AND AWARDS COMMITTEE**

### **SUPPLEMENTAL/BID BULLETIN**

#### **Addendum No. 1**

(Procurement of Various Laboratory Equipment for CY 2018)

This Addendum No. 1 dated June 7, 2018 is issued to amend items/articles in the Invitation to Bid, Bid Data Sheet, Special Conditions of Contract, Section VI. Schedule of Requirements – CERTIFICATION and Technical Specifications for the Procurement of Various Laboratory Equipment for CY 2018 on June 18, 2018. This shall form an integral part of the Bid Documents.

### **Invitation to Bid**

#### **ORIGINAL:**

The *Western Visayas Medical Center*, through Hospital Generated Income, in the amount of ***Thirteen Million Seven Hundred Eighty One Pesos Only (Php13,781,265.00)*** being the Approved Budget for the Contract (ABC) to payments under the contract for the ***Procurement of Various Laboratory Equipment for CY 2018***. Bids received in excess of the ABC shall be automatically rejected at bid opening.

#### **AMENDED:**

The *Western Visayas Medical Center*, through Hospital Generated Income, in the amount of ***Thirteen Million Six Hundred Seventy Nine Thousand One Hundred Forty Five Pesos Only (Php13,679,145.00)*** being the Approved Budget for the Contract (ABC) to payments under the contract for the ***Procurement of Various Laboratory Equipment for CY 2018***. Bids received in excess of the ABC shall be automatically rejected at bid opening.

All other articles in the **Invitation to Bid** remain the same.

## Bid Data Sheet

<i><b>ORIGINAL</b></i>	<i><b>AMENDED</b></i>
<p>2 The Funding Source is:</p> <p>The Government of the Philippines (GOP) through the Hospital Generated Income, in the amount of <i><b>Thirteen Million Seven Hundred Eighty One Pesos Only (Php13,781,265.00).</b></i></p> <p>The name of the Project is: <i><b>Procurement of Various Laboratory Equipment for CY 2018.</b></i></p>	<p>2 The Funding Source is:</p> <p>The Government of the Philippines (GOP) through the Hospital Generated Income, in the amount of <u><i><b>Thirteen Million Six Hundred Seventy Nine Thousand One Hundred Forty Five Pesos Only (Php13,679,145.00).</b></i></u></p> <p>The name of the Project is: <i><b>Procurement of Various Laboratory Equipment for CY 2018.</b></i></p>
<p>13.2 The ABC is <i><b>Thirteen Million Seven Hundred Eighty One Pesos Only (Php13,781,265.00).</b></i> Any bid with a financial component exceeding this amount shall not be accepted.</p>	<p>13.2 The ABC is <u><i><b>Thirteen Million Six Hundred Seventy Nine Thousand One Hundred Forty Five Pesos Only (Php13,679,145.00).</b></i></u> Any bid with a financial component exceeding this amount shall not be accepted.</p>
<p>18.1 The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <ol style="list-style-type: none"> <li>1. The amount of not less than <u><i><b>₱275,625.30 (2% of ABC)</b></i></u> if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</li> <li>2. The amount of not less than <u><i><b>₱689,063.25 (5% of ABC)</b></i></u> if bid security is in Surety Bond.</li> </ol>	<p>18.1 The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <ol style="list-style-type: none"> <li>1. The amount of not less than <u><i><b>₱273,582.90 (2% of ABC)</b></i></u> if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</li> <li>2. The amount of not less than <u><i><b>₱683,957.25 (5% of ABC)</b></i></u> if bid security is in Surety Bond.</li> </ol>
<p style="text-align: center;">All other articles in the <b>Bid Data Sheet</b> remain the same.</p>	

## Special Conditions of Contract

<b><i>ORIGINAL</i></b>	<b><i>AMENDED</i></b>
<p>1.1(j) The Funding Source is:</p> <p>The Government of the Philippines (GOP) through the Hospital Generated Income, in the amount of <b><i>Thirteen Million Seven Hundred Eighty One Pesos Only (Php13,781,265.00).</i></b></p>	<p>1.1(j) The Funding Source is:</p> <p>The Government of the Philippines (GOP) through the Hospital Generated Income, in the amount of <b><u>Thirteen Million Six Hundred Seventy Nine Thousand One Hundred Forty Five Pesos Only (Php13,679,145.00).</u></b></p>
<p>6.2 <b><u>Scope of the Contract</u></b></p> <ul style="list-style-type: none"> <li>• <i>Contract for the Laboratory Equipment is one year subject to the quarterly evaluation of the Property and Supply Office based on the following criteria; (a) The Quality of goods, and the (b) Efficiency in the delivery of goods.</i></li> <li>• <i>Warranty on parts and services for Equipment is two (2) years free services, 1 year free parts and free technical support until the contract ends.</i></li> <li>• <i>Brand new unit, with Onsite calibration and validation with certificate of conformity.</i></li> <li>• <i>With available technical support personnel for corrective and preventive maintenance within 24-48 hours.</i></li> <li>• <i>Machines / analyzers must be provided with Uninterrupted Power Supply compatible with the machine and with power surge protection (AVR/UPS).</i></li> <li>• <i>Provides Operation and maintenance training for End-Users with certificate.</i></li> </ul>	<p>6.2 <b><u>Scope of the Contract</u></b></p> <ul style="list-style-type: none"> <li>• <i>Contract for the Laboratory Equipment is one year subject to the quarterly evaluation of the Property and Supply Office based on the following criteria; (a) The Quality of goods, and the (b) Efficiency in the delivery of goods.</i></li> <li>• <i>Warranty on parts and services for Equipment is two (2) years free services, 1 year free parts and free technical support until the contract ends.</i></li> <li>• <i>Brand new unit, with Onsite calibration and validation with certificate of conformity.</i></li> <li>• <i>With available technical support personnel for corrective and preventive maintenance within 24-48 hours.</i></li> <li>• <i>Machines / analyzers must be provided with <b><u>1.5-3 kVA</u></b> Uninterrupted Power Supply compatible with the machine and with power surge protection (AVR/UPS).</i></li> <li>• <i>Provides Operation and maintenance training for End-Users with certificate.</i></li> </ul>
<p>All other terms and conditions in <b>Special Conditions of Contract</b> remain the same.</p>	

## Section VI. Schedule of Requirements, CERTIFICATION

### ORIGINAL:

#### CERTIFICATION

This is to certify that the delivery of the following item/s that I have bid and found to be the lowest calculated and responsive bid is accordance to the agency's specification shall be delivered to the Western Visayas Medical Center within Thirty (30) Days from receipt of the Notice to Proceed:

Item #	Description	Quantity	Unit	Brand/Country of Origin	Delivery Period
					As per Schedule of Delivery of the End User

This certification is issued as requirement for the June 18, 2018 *Procurement of Various Laboratory Equipment for CY 2018* for the use of the Western Visayas Medical Center.

Issued this 18<sup>th</sup> day of June, 2018.

Agency's Proprietor/Manager/President or  
Authorized Representative

### AMENDED:

#### CERTIFICATION

This is to certify that the delivery of the following item/s that I have bid and found to be the lowest calculated and responsive bid is accordance to the agency's specification shall be delivered to the Western Visayas Medical Center within Sixty (60) Days from receipt of the Notice to Proceed:

Item #	Description	Quantity	Unit	Brand/Country of Origin	Delivery Period
					As per Schedule of Delivery of the End User

This certification is issued as requirement for the June 18, 2018 *Procurement of Various Laboratory Equipment for CY 2018* for the use of the Western Visayas Medical Center.

Issued this 18<sup>th</sup> day of June, 2018.

Agency's Proprietor/Manager/President or  
Authorized Representative

## Technical Specifications

**ORIGINAL:**

Item #	Item/Description	Unit	Quantity	ABC/ Unit	Total ABC	Statement of Compliance
3	<b>Fume hood</b> <b>Specifications:</b> Inner Dimensions (WxDxH): 1050mm x 650mm x 750 mm Outer Dimensions (WxDxH): 1200 mm x 750 mm x 2250 mm Air Volume (Approx): 1450 CMH (+/- 10%) -Ducted -Designed for acid digestion -Interior material: Stainless steel -Sash door: Tempered glass sliding door - Utility device: With water faucet -Air Volume: approximately 1450 m <sup>3</sup> /hour (+/- 10%) -With Incandescent lamp -with built-in Fan	Unit	1	385,930.00	385,930.00	
5	<b>Vacuum Pump</b> <b>Specifications:</b> Motor: 0.5 Horse Power Max. Pressure: 60 PSI Stage: 1	Unit	2	20,000.00	40,000.00	
6	<b>Water Bath, Digital,10 liters capacity w/ Certificate of Calibration</b> <b>Specifications:</b> <ul style="list-style-type: none"> <li>- Stainless steel: interior and casing</li> <li>- Volume: 10 liters capacity</li> <li>- Temperature range: ambient temperature up to 100 °C</li> <li>- Dimension: wxhxd : 356 x238 x 578</li> </ul>	Unit	1	54,611.00	54,611.00	
7	<b>Water Purification System, Ultrapure Water</b> <b>Specifications:</b> <ul style="list-style-type: none"> <li>- High pure water</li> <li>- For ultra trace analysis</li> <li>- Product range: 30L/hour</li> <li>- Power consumption : 80W</li> <li>- Dimension: L X W X H: 390 x 540 x 500mm</li> <li>- Weight (kg): 29±1</li> </ul>	Unit	1	195,000.00	195,000.00	
12	<b>Pipettor 0.2 - 20uL with 0.1uL increments</b>	Unit	1	10,000.00	10,000.00	
14	<b>30 PLACER MECHANICAL CENTRIFUGE COMPATIBLE W/ FIXED ANGLE ROTOR</b> <b>Specifications:</b> Unlimited Running Time Speed: 300rpm-15200rpm Certifications/Compliance: IVD Compliant, IEC 61010-1, IEC 61010-2-202, IEC 61010-2-101. UL listed, CSA certified, CE marked, Certified Biosafety. Auto-lock Rotor System Motorized Lid latch Imbalance detection system uses SmartSpin Technology Easy adjust speed and time even during runtime	Unit	3	250,000.00	750,000.00	

	Dedicated program keys					
19	<b>Loop Incinerator</b> <b>Specifications:</b> a. sterilize metal inoculating loops and needles without using an open flame b. Infrared heat inside a ceramic tube c. Weighed base for stability d. Operates on 120V 60Hz 220V	Unit	2	19,000.00	38,000.00	
25	<b>Colorimeter, Chloramine with test reagents</b>	Unit	1	44,120.00	44,120.00	
26	<b>Colorimeter, Chlorine (Free and Total) with test Reagents</b>	Unit	1	38,000.00	38,000.00	

**AMENDED:**

Item #	Item/Description	Unit	Quantity	ABC/ Unit	Total ABC	Statement of Compliance
3	<b>Fume hood</b> <b>Specifications:</b> Inner Dimensions (WxDxH): 1050mm x 650mm x 750 mm <u><b>±60 mm</b></u> Outer Dimensions (WxDxH):1200 mm x750 mm x 2250 mm <u><b>±60 mm</b></u> Air Volume (Approx): 1450 CMH (+/- 10%) -Ducted -Designed for acid digestion -Interior material: Stainless steel -Sash door: Tempered glass sliding door - Utility device: With water faucet -Air Volume: approximately 1450 m <sup>3</sup> /hour (+/- 10%) -With Incandescent lamp -with built-in Fan	Unit	1	385,930.00	385,930.00	
5	<b>Vacuum Pump</b> <b>Specifications:</b> Motor: 0.5 Horse Power Max. Pressure: 60 PSI Stage: 1	Unit	<u><b>1</b></u>	<u><b>20,000.00</b></u>	<u><b>20,000.00</b></u>	
6	<b>Water Bath, Digital,10 liters capacity w/ Certificate of Calibration</b> <b>Specifications:</b> <ul style="list-style-type: none"> <li>Stainless steel: interior and casing</li> <li>Volume: 10 liters capacity</li> <li>Temperature range: ambient temperature up to 100 °C</li> <li>Dimension: wxhxd : 356 x238 x 578<u><b>mm ± 60 mm</b></u></li> </ul>	Unit	1	54,611.00	54,611.00	
7	<b>Water Purification System, Ultrapure Water</b> <b>Specifications:</b> <ul style="list-style-type: none"> <li>High pure water</li> <li>For ultra trace analysis</li> <li>Product range: 30L/hour</li> <li>Power consumption : 80W</li> <li>Dimension: LxWxH: 390 x 540 x 500mm <u><b>±60mm</b></u></li> </ul>	Unit	1	195,000.00	195,000.00	

	• Weight (kg): 29±1					
12	<b>Pipettor 0.2 - 20uL with .01uL increments</b>	Unit	1	10,000.00	10,000.00	
14	<b>30 PLACER MECHANICAL CENTRIFUGE COMPATIBLE W/ FIXED ANGLE ROTOR</b>  Specifications: <u>Fits 13x75mm or 13x 100 mm round bottom Test tubes or 4-6 ml capacity test tubes.</u> Unlimited Running Time Speed: 300rpm-15200rpm Certifications/Compliance: IVD Compliant, IEC 61010-1, IEC 61010-2-202, IEC 61010-2-101. UL listed, CSA certified, CE marked, Certified Biosafety. Auto-lock Rotor System Motorized Lid latch Imbalance detection system uses Smart Spin Technology Easy adjust speed and time even during runtime Dedicated program keys	Unit	3	250,000.00	750,000.00	
19	<b>Loop Incinerator</b>  <b>Specifications:</b> a. sterilize metal inoculating loops and needles without using an open flame b. Infrared heat inside a ceramic tube c. Weighed base for stability <u>d. Operates on 60Hz 220V</u>	Unit	2	19,000.00	38,000.00	
<del>25</del>	<del>Colorimeter, Chloramine with test reagents</del>	<del>Unit</del>	<del>1</del>	<del>44,120.00</del>	<del>44,120.00</del>	
<del>26</del>	<del>Colorimeter, Chlorine (Free and Total) with test Reagents</del>	<del>Unit</del>	<del>1</del>	<del>38,000.00</del>	<del>38,000.00</del>	
All other items in the <b>Technical Specifications</b> remain the same.						

For guidance and information of all concerned:

SGD. ERIC J. MONTERO, MD  
 Chairman, Bids and Awards Committee